

Cat No. RCDH101 Lot No. A01RG03**MFG date 03 Aug 2018**

*Store at -20°C, Performance guaranteed until Exp. Date

Expiration date Aug 2021■ **Properties and biological tests**

Appearance	Transparent orange liquid
Mycoplasma ⁽¹⁾	Negative
Sterility test ⁽²⁾	Pass
pH ⁽³⁾	7.41
Osmotic pressure ⁽⁴⁾	333 mOsm/kg

⁽¹⁾Mycoplasma detection performed by PCR analysis.

⁽²⁾Sterility was confirmed by visual inspection after culturing for a week.

⁽³⁾Measured after 12 hours incubation under conditions of 37°C and 5% CO₂.

⁽⁴⁾Measured using the freezing point depression method.