

Cat No. RCHEMD001 Lot No. R1M151215**MFG date 15 Dec 2015**

*Store at -20°C, Performance guaranteed until Exp. Date

Expiration date Dec 2018**■ Properties and biological tests**

Appearance	Transparent orange liquid
Aseptic test ⁽¹⁾	Negative (Day 14)
Mycoplasma ⁽²⁾	Negative
pH ⁽³⁾	7.30
Endotoxin	0.032 EU/mL
Osmotic pressure ⁽⁴⁾	332 mOsm/kg

⁽¹⁾Cultured using thioglycolic acid medium for aseptic test and SCD medium.

⁽²⁾Mycoplasma detection performed by PCR analysis.

⁽³⁾Measured after 12 hours incubation under conditions of 37°C and 5% CO₂.

⁽⁴⁾Measured using the freezing point depression method.

■ Culture test performed using human iPS cells**Methods**


Human iPS cells (201B7; Takahashi, K. et al., *Cell*, 131, 861-72, 2007) were cultured in the Primate ES cell medium in the presence of MEF feeder cells for 4 days.

After 4 passages, the cell morphology was observed.

Alkaline phosphatase activity was ascertained using the ALP substrate kit III (VECTOR).

Results

Alkaline phosphatase activity: Qualified



Scale: 250 µm

