

Cat No. RCHEFM001 Lot No. B01QN09**MFG date 9 Nov 2017**

*Store at -20°C, Performance guaranteed until Exp. Date

Expiration date Nov 2020■ **Properties and biological tests**

Appearance	Transparent orange liquid
Aseptic test ⁽¹⁾	Negative (Day 14)
Mycoplasma ⁽²⁾	Negative
pH ⁽³⁾	7.89
Endotoxin	0.017EU/mL

⁽¹⁾Cultured using thioglycolic acid medium for aseptic test and SCD medium.

⁽²⁾Mycoplasma detection performed by PCR analysis.

⁽³⁾Measured after 12 hours incubation under conditions of 37°C and 5% CO₂.